

**OFFICE OF THE DIRECTOR CUM PRINCIPAL
ATAL BIHARI VAJPAYEE GOVERNMENT INSTITUTE OF ENGINEERING &
TECHNOLOGY,
PRAGATINAGAR, DISTRICT-SHIMLA (HIMACHAL PRADESH) 171202.**

Tele: 01783-253918 Fax: 253919 Email: director.abv@gmail.com Website: <https://www.abvgiet.ac.in/>
No:ABVGIET/Store/B.T./Limited Tender/Civil/-2026-27- 160 - 168 Date: 22/01/2026

LIMITED TENDER NOTICE

(Limited Tender No. ABVGIET/Store/B.Tech/Limited Tender/2026-27)

Sealed limited tenders are invited from eligible manufacturers/firms/authorized dealers for the supply of Civil Engineering Laboratory Equipment as per specifications given in this tender document. The bids must be accompanied with Earnest Money as prescribed in the bidding schedule in the form of Demand Draft duly pledged in favour of Director-cum-Principal, ABVGIET Pragatinagar.

A complete set of bid documents along with detailed terms and conditions can be obtained from the institute website (abvgiet.ac.in) or from the office of the undersigned.

The following procedure will be adopted:

1. Tenders in a sealed cover shall comprise two separate envelops, each envelop shall contain separately the technical bid and the financial bid. The envelops shall be marked as '**Technical Bid**' and '**Financial Bid**' in bold and legible letters to avoid any confusion.
2. Initially technical bid will be opened.
3. The technical bid shall be evaluated without reference to the price and will be rejected which does not conform to the specific technical criteria.
4. During the technical evaluation, no amendments to the technical proposal shall be permitted.
5. The financial bid will be opened after the evaluation of technically qualified bidder.
6. The bid found to be the lowest evaluated in cumulative shall be accepted.
7. This office will not be held responsible for the postal delay, if any.
8. Undersigned reserves the right to accept or reject all or any of the tenders without assigning any reason(s).

h sm

Director cum Principal
Atal Bihari Vajpayee
Govt. Institute of Engg. & Technology
Pragatinagar Distt. Shimla HP 171202.
M

→ website or f w.
h

**OFFICE OF THE DIRECTOR CUM PRINCIPAL
ATAL BIHARI VAJPAYEE GOVERNMENT INSTITUTE OF ENGINEERING &
TECHNOLOGY,
PRAGATINAGAR, DISTRICT-SHIMLA (HIMACHAL PRADESH) 171202.**

Tele: 01783-253918 Fax: 253919 Email: director.abv@gmail.com Website: <https://www.abvgiet.ac.in/>

No. ABVGIET/STR/ Limited Tender / 2026-27- 160 - 168 S. No.....

**Limited Tender Document for the supply of Machinery and Equipment's required in ABVGIET
Pragatinagar, Distt. Shimla-171202 (H.P)**

General Information

1	Limited tender for supply of	Machinery & Equipment
2	Limited Tender Reference	ABVGIET/Store/B.Tech/Limited Tender/2026-27/01
3	Last date and time for receipt of tender	06-02-2026 at 02:00 PM
4	Time and date of opening of tender	06-02-2026 at 03:00 PM
5	Place of opening of tender:	Office of Director-cum-Principal, ABVGIET Pragatinagar
6	Address for communication	ABVGIET Pragatinagar, Distt. Shimla-171202 (H.P)
7	Estimated cost	9,80,000/- Rupees only
8	Earnest money	Rupees 10,000/- only
9	Performance Security	Rupees 25,000/- only
10	Cost of limited tender document	Rupees 500/- only (Rs. 600/- by post)

L. S. M.

Director cum Principal
Atal Bihari Vajpayee
Govt. Institute of Engg. & Technology
Pragatinagar Distt. Shimla HP 171202.

SECTION-A

1. Name of the firm/Organization : -----
2. (a) Registration No : -----
(b) GST No.:
(c) TAN/PAN :
3. Complete Postal address : -----
: -----
: -----
4. Phone No. with STD code : -----
5. E-mail address (if any) : -----
6. Whether Manufacturer/Dealer/Supplier : -----
7. Total Amount of items Tendered
(Excluding taxes) in words : -----


8. Detail of earnest Money Draft No. & Date : -----
Amount : -----
Bank : -----

(Signature of the Tenderer with seal)

SECTION-B

Schedule of Requirement

Sr. No.	Description of item(s)	Qty.
Schedule – A		
	Supply of Machinery and Equipment required in ABVGIET Pragatinagar	As per list attached in section C



Director cum Principal
Atal Bihari Vajpayee
Govt. Institute of Engg. & Technology
Pragatinagar Distt. Shimla HP 171202

SECTION-C

DETAILED SPECIFICATION OF THE ITEMS MENTIONED IN SCHEDULE-A

S. N	Description of item	Qty.
LIST of Machinery and Equipment required in ABVGIET Pragatinagar (HP)		
1	<p>Chemical Oxygen Demand (COD) Portable Photometer</p> <p>Measurement Ranges & Accuracy</p> <ul style="list-style-type: none"> • Low Range (LR): 0 to 150 mg/L (as O₂) <ul style="list-style-type: none"> ◦ Resolution: 1 mg/L ◦ Accuracy: ±5 mg/L or ±4% of reading at 25°C, whichever is greater ◦ LED Wavelength: 420 nm • Medium Range (MR): 0 to 1500 mg/L (as O₂) <ul style="list-style-type: none"> ◦ Resolution: 1 mg/L ◦ Accuracy: ±15 mg/L or ±4% of reading at 25°C ◦ LED Wavelength: 610 nm • High Range (HR): 0 to 15000 mg/L (as O₂) <ul style="list-style-type: none"> ◦ Resolution: 1 mg/L ◦ Accuracy: ±150 mg/L or ±2% of reading at 25°C ◦ LED Wavelength: 610 nm • Ultra High Range (UHR): 0 to 60.0 g/L (as O₂) <ul style="list-style-type: none"> ◦ Resolution: 0.1 g/L ◦ Accuracy: ±0.5 g/L ±3% of reading at 25°C ◦ LED Wavelength: 610 nm <p>Measurement System</p> <ul style="list-style-type: none"> • Light Source: LED • Bandpass Filter: Wavelengths at 420 nm & 610 nm, bandwidth ≈ 8 nm, accuracy ±1.0 nm • Light Detector: Silicon photocell • Cuvette Type: Round, 16 mm diameter <p>Photometer Features</p> <ul style="list-style-type: none"> • Data logging for up to 200 readings • Display: 128 x 64 pixel backlit B/W LCD • Auto-off: After 15 minutes of inactivity (extended to 30 minutes if zeroed) • Battery Type: 1.5 V AA alkaline (3 units) • Battery Life: Over 10 000 measurements (without backlight) 	1

	<p>Environmental & Physical</p> <ul style="list-style-type: none"> • Operating Environment: 0 to 50°C (32 to 122°F); 0 to 100% relative humidity (non-serviceable) • Dimensions: 142.5 x 102.5 x 50.5 mm (5.6 x 4.0 x 2.0 in.) • Weight: 380 g (with batteries) • Protection Rating: IP67 waterproof, floating casing <p>Method</p> <ul style="list-style-type: none"> • Based on an adaptation of the US EPA Method 410.4 for COD determination in surface waters and wastewater samples. <p>Accessories Included</p> <ul style="list-style-type: none"> • Adapter for 16 mm cuvette vial • Batteries (3 x AA) • Quick reference guide and instrument quality certificate <p>Note- Equipment should be supplied with minimum two year warranty and there should be enough reagents to conduct tests for atleast 100 samples of water</p>	
2	<p>Handheld Multi Parameter Water Quality Analyzer</p> <p>Core Features</p> <p>Waterproof housing (meter IP67; probe IP68 continuous immersion).</p> <p>Graphic backlit LCD display with on-screen help and trend graphing.</p> <p>Automatic recognition of connected sensors.</p> <p>Data logging with time/date and optional GPS coordinate recording.</p> <p>USB connectivity for PC data transfer.</p> <p>Fast Tracker™ Tag Identification System for site labeling.</p> <p>Automatic temperature and barometric pressure compensation (for dissolved oxygen).</p> <p>Measured Parameters & Ranges</p> <p>pH and Related</p> <p>Range: 0.00 to 14.00 pH</p> <p>Resolution: 0.01 pH</p>	1

	<p>Accuracy: ± 0.02 pH</p> <p>Calibration: Automatic 1–3 points; multiple standard buffers supported</p> <p>mV / ORP</p> <p>mV Range: ± 600.0 mV; resolution 0.1 mV; accuracy ± 0.5 mV</p> <p>ORP Range: ± 2000 mV; resolution 0.1 mV; accuracy ± 1.0 mV</p> <p>Conductivity</p> <p>Range: 0 to 200 mS/cm (absolute up to 400 mS/cm)</p> <p>Resolution and auto-range depending on range selected</p> <p>Accuracy: $\pm 1.0\%$ of reading or $\pm 1 \mu\text{S}/\text{cm}$, whichever greater</p> <p>Calibration: Automatic single-point using multiple standard solutions</p> <p>TDS</p> <p>Range: 0 to 400,000 ppm (mg/L) (depends on factor)</p> <p>Resolution varies by range and units selected</p> <p>Accuracy: $\pm 1\%$ of reading or ± 1 ppm, whichever greater</p> <p>Resistivity</p> <p>Range: 0 to 999,999 $\Omega \cdot \text{cm}$; up to 1.0000 $\text{M}\Omega \cdot \text{cm}$</p> <p>Salinity / Seawater</p> <p>Salinity: 0.00 to 70.00 PSU</p> <p>Seawater conductivity parameters (σ, σ_0, σ_{15}) included</p> <p>Turbidity (if equipped)</p> <p>Range: 0.0 to 99.9 FNU and 100 to 1000 FNU</p> <p>Resolution: 0.1 FNU for 0–99.9; 1 FNU above</p> <p>Dissolved Oxygen</p> <p>Range: 0.0 to 50.00 ppm (mg/L) and 0.0 to 500.0%</p>	
--	---	--

	<p>Resolution: 0.01 ppm and 0.1%</p> <p>Environmental & Physical</p> <p>Operating Temperature: Typically 0 to 50°C</p> <p>Meter Dimensions: ~221 × 115 × 55 mm</p> <p>Weight: ~750 g (meter only)</p> <p>Power Supply: Four 1.5 V cells (alkaline or NiMH), with optional 12 V adaptor</p> <p>Maximum probe immersion depth: 20 m (approx.)</p> <p>Logging & Connectivity</p> <p>Data Logging: Up to 44,000 samples from meter</p> <p>Logging Intervals: 1 second to 3 hours</p> <p>GPS: Optional 12-channel receiver for location tracking (meters with GPS option)</p> <p>PC Interface: USB connection with software for download and analysis</p> <p>Calibration & User Aids</p> <p>Quick calibration routines for pH, conductivity, and dissolved oxygen</p> <p>Contextual on-screen help</p> <p>GLP data storage (calibration records and expiration info)</p> <p>Note- Equipment should be supplied with minimum two year warranty and there should be enough reagents to conduct tests for at least 100 samples of water</p>	
3	<p>Portable Dissolved Oxygen Meter</p> <p>Measurement</p> <ul style="list-style-type: none"> Dissolved oxygen range: 0.0 to 19.9 mg/L (ppm) Resolution (DO): 0.1 mg/L (ppm) Accuracy (DO): ±1.5% of full scale Calibration: Manual calibration (one or two points – zero and slope) Temperature compensation: Automatic <p>Temperature</p> <ul style="list-style-type: none"> Measuring range: -5.0 to 50.0°C (32.0 to 122.0°F) 	1

		<ul style="list-style-type: none"> • Resolution: 0.1°C / 0.1°F • Accuracy (temperature): $\pm 0.2^\circ\text{C}$ ($\pm 1^\circ\text{F}$) excluding probe error <p>Probe</p> <ul style="list-style-type: none"> • Sensor type: Polarographic dissolved oxygen probe • Internal temperature sensor included • Cable length (standard): 4 m (13 ft) <p>Environmental & Physical</p> <ul style="list-style-type: none"> • Operating temperature: 0 to 50°C (32 to 122°F) • Relative humidity: up to 95% • Waterproof/rugged housing designed for field use • Dimensions: 185 × 72 × 36 mm (approx.) • Weight: ≈ 300 g <p>Power</p> <ul style="list-style-type: none"> • Battery type: 3 × 1.5 V batteries (AAA/AA depending on source) • Battery life: $\sim 1,000$ hours continuous use <p>Probe Accessories / Included Items</p> <ul style="list-style-type: none"> • Do probe with cable • Replaceable PTFE membrane caps • Electrolyte fill solution <p>Protective shield and rugged carrying case</p> <p>Note- Equipment should be supplied with minimum two year warranty and there should be enough reagents to conduct tests for attest 100 samples of water</p>	
4	Waterproof Portable EC/TDS/Resistivity/Salinity Meter	<p>Measured Parameters & Ranges</p> <p>Conductivity (EC)</p> <ul style="list-style-type: none"> • Range (temperature-compensated): 0.000 to 9.999 $\mu\text{S}/\text{cm}$; 10.00 to 99.99 $\mu\text{S}/\text{cm}$; 100.0 to 999.9 $\mu\text{S}/\text{cm}$; 1.000 to 9.999 mS/cm; 10.00 to 99.99 mS/cm; 100.0 to 1000.0 mS/cm (<i>actual conductivity; temp-compensated to 400 mS/cm</i>) • Resolution: 0.001 $\mu\text{S}/\text{cm}$–0.1 mS/cm (range dependent) • Accuracy: $\pm 1\%$ of reading (± 0.01 $\mu\text{S}/\text{cm}$ or 1 digit, whichever greater) • Calibration: Automatic up to 5 points with 7 memorized standards (0.00 $\mu\text{S}/\text{cm}$, 84.0 $\mu\text{S}/\text{cm}$, 1.413 mS/cm, 5.00 mS/cm, 12.88 mS/cm, 80.0 mS/cm, 111.8 mS/cm) 	1

	<p>Total Dissolved Solids (TDS)</p> <ul style="list-style-type: none"> • Range: 0.00 to 99.99 ppm; 100.0 to 999.9 ppm; 1.000 to 9.999 g/L; 10.00 to 99.99 g/L; 100.0 to 400.0 g/L • Resolution: 0.01 ppm to 0.1 g/L (range dependent) • Accuracy: $\pm 1\%$ of reading (± 0.05 mg/L or 1 digit) • TDS Factor: 0.40 to 1.00 selectable <p>Resistivity</p> <ul style="list-style-type: none"> • Range: 1.0 to 99.9 $\Omega \cdot \text{cm}$; 100 to 999 $\Omega \cdot \text{cm}$; 1.00 to 9.99 $\text{K}\Omega \cdot \text{cm}$; 10.0 to 99.9 $\text{K}\Omega \cdot \text{cm}$; 100 to 999 $\text{K}\Omega \cdot \text{cm}$; 1.00 to 9.99 $\text{M}\Omega \cdot \text{cm}$; 10.0 to 100.0 $\text{M}\Omega \cdot \text{cm}$ • Resolution: 0.1 $\Omega \cdot \text{cm}$ to 0.1 $\text{M}\Omega \cdot \text{cm}$ (range dependent) • Accuracy: $\pm 1\%$ of reading ($\pm 10 \Omega$ or 1 digit) <p>Salinity</p> <ul style="list-style-type: none"> • Scales: % NaCl (0.0 to 400.0%); Practical Salinity (0.00 to 42.00 PSU); Natural Seawater (0.00 to 80.00 ppt) • Resolution: 0.1%; 0.01 • Accuracy: $\pm 1\%$ of reading • Calibration: Single-point in % range (with appropriate standard); conductivity calibration used for other scales <p>Temperature</p> <ul style="list-style-type: none"> • Range: -20.0 to 120.0°C (-4.0 to 248.0°F) • Resolution: 0.1°C / 0.1°F • Accuracy: $\pm 0.2^\circ\text{C}$; $\pm 0.4^\circ\text{F}$ (excluding probe error) • Calibration: One or two points • Compensation: No TC, Linear, or Non-Linear (ISO/DIN 7888) <p>Probe & Measurement Details</p> <ul style="list-style-type: none"> • Electrode/Probe: Four-ring conductivity/TDS probe with internal temperature sensor (typically supplied) • Probe Cable: ~4 m (13.2 ft) • Cell Constant: Adjustable per probe type recognition • Measurement Modes: Auto range, auto end, lock, fixed range • Reference Temperatures: 15°C, 20°C, 25°C • Temperature Compensation Coefficient: 0.00 to 10.00%/$^\circ\text{C}$ <p>Data Logging & Connectivity</p> <ul style="list-style-type: none"> • Data Logging: <ul style="list-style-type: none"> • Log-on-demand: up to 400 samples 	
--	---	--

	<ul style="list-style-type: none"> • Lot logging (interval logging): up to 1000 samples at selectable intervals (5 sec – 180 min) • Profiles: Up to 10 memorized profiles • PC Connectivity: Opto-isolated USB interface with software and micro USB cable support <p>Environmental & Physical</p> <ul style="list-style-type: none"> • Environment: 0 to 50°C (32 to 122°F); 100% relative humidity; waterproof rating IP67 • Dimensions: ~185 × 93 × 35.2 mm • Weight: ~400 g • Power: 4 × 1.5V AA batteries • Battery Life: ~100 hours (without backlight); ~25 hours (with backlight) • Auto-Off: User selectable or disabled <p>Calibration Standards (Typical)</p> <ul style="list-style-type: none"> • Conductivity: 0.00, 84.0 μS/cm, 1.413, 5.00, 12.88, 80.0, 111.8 mS/cm <p>Salinity: Single point with % standard</p> <p>Note- Equipment should be supplied with minimum two year warranty and there should be enough reagents to conduct tests for attest 100 samples of water</p>	
5	<p>Portable Photometer</p> <p>Measurement Parameter</p> <ul style="list-style-type: none"> • Color of Water (PCU): 0 to 500 PCU • Resolution: 1 PCU • Accuracy: ± 10 PCU $\pm 5\%$ of reading at 25 °C • Measurement Method: Adaptation of the Standard Methods for the Examination of Water and Wastewater, Colorimetric Platinum Cobalt Method <p>Optical & Measurement System</p> <ul style="list-style-type: none"> • Light Source: Light-emitting diode (LED) with narrow band interference filter (~420 nm) • Bandpass Filter Bandwidth: ~8 nm; wavelength accuracy ± 1.0 nm • Light Detector: Silicon photocell • Cuvette Type: Round, 24.6 mm diameter (22 mm inside) <p>Display & Controls</p> <ul style="list-style-type: none"> • Display: 128 × 64 pixel backlit B/W LCD 	1

	<ul style="list-style-type: none"> • User Interface: Virtual keys with tutorial mode guidance • Logging: Automatic logging of last 50 readings <p>Power & Battery Battery Type: 1.5 V AA alkaline (3 units)</p> <ul style="list-style-type: none"> • Battery Life: >800 measurements (without backlight) • Automatic Shut-off: Activates after ~15 minutes of inactivity (longer before a measurement) <p>Environmental & Physical</p> <ul style="list-style-type: none"> • Operating Environment: 0 to 50°C (32 to 122°F); 0 to 100% relative humidity (non-serviceable) • Protection: Waterproof (typically IP67) with floating case design • Dimensions: ~142.5 × 102.5 × 50.5 mm (5.6 × 4.0 × 2.0 in.) • Weight: ~380 g (with batteries) <p>Additional Functionality</p> <ul style="list-style-type: none"> • CAL Check Feature: For performance verification and calibration • Tutorial Mode: Step-by-step graphical guidance during setup/measurement • Display Features: GLP data, battery status, error messages (e.g., no cap) <p>Typical Supply (Meter-Only)</p> <ul style="list-style-type: none"> • Sample cuvettes (2) • Sample caps (2) • Plastic stoppers (2) • Alkaline batteries (3) • Instruction manual <p>Instrument quality certificate</p> <p>Note- Equipment should be supplied with minimum two year warranty and there should be enough reagents to conduct tests for attest 100 samples of water</p>	
--	---	--

For (Authorized Signatory)
 Name of the firm.....
 Complete address.....

Seal Stamp.....



Director cum Principal

Atal Bihari Vajpayee Govt. Institute of Engg. & Technology

Pragatinagar Distt. Shimla HP 171202

Page 14 of 18

SECTION-D
PRICE SCHEDULE

S.N	Description of item	Qty	Cost per unit (Rupees).	Total Cost (Rupees). (GST included)
1	Chemical Oxygen Demand (COD) Portable Photometer	1		
2	Handheld Multi Parameter Water Quality Analyzer	1		
3	Portable Dissolved Oxygen Meter	1		
4	Waterproof Portable EC/TDS/Resistivity/Salinity Meter	1		
5	Portable Photometer	1		

Total Bid Price (Excluding Taxes in Rs) _____
 (In words) _____
 Date _____
 Place _____

(Signature of the bidder with seal)
 Name and Address-----

TERMS AND CONDITIONS

1. Bidders are required to submit the attested copy of Registration (latest) clearly mentioning the period of validity. Any firm failing to submit the same will not be entertained.
2. Rate has to be quoted for complete set (including all sub-parts/accessories, if any) of machinery/equipment to be supplied. Bidder quoting the lowest rate for complete set shall only be considered fulfilling all other requisite conditions.
3. Bidder quoting the lowest amount for complete tender shall be considered qualified, fulfilling all other requisite conditions.
4. The tenderer shall submit the latest income tax return statement. This will help in assessing the capability of the bidder to supply the items mentioned in the schedule of bid document.
5. The bidder shall also submit the proof that it is GST payee. The attested photocopy of the same is required to be submitted along with the bid document.
6. The rates quoted should be F.O.R. destination at ABVGIET Pragatinagar, Distt. Shimla-171202 (H.P). The rates must also include installation/Demonstration and Training expenses, if any.
7. Standard GST rates are applicable, if any.
8. The duly constituted College Committee shall inspect the material to satisfy itself for verifying specifications as mentioned in bid document.
9. The supplier shall be required to complete the delivery at college (including inspection) within stipulated days as mentioned in the supply order.
10. The items/equipments shall remain under Guarantee/Warranty by the supplier.
11. The Guaranty/Warranty shall be effective from the actual date of installation in case of items of machinery and equipment.
12. 100% payment will be made within one month against physical delivery of inspected/accepted stores duly supported with satisfactory inspection note and after receipt of correct goods at consignee site/destination.
13. Successful bidder has to give the Bank Account Detail such as account name, number, IFSC code of bank, bank name, branch and PAN. Also provide AADHAR number in case the account is on the name of a person. Payment will be directly paid to account holder by Govt. treasury after completing all the required formalities.
14. Earnest money as mentioned in the document shall be required to be paid by the tenderer along with the tender document in shape of account payee demand draft or duly pledged fixed deposit receipts in favor of the Director-cum-Principal, ABVGIET Pragatinagar, Distt. Shimla-171202 (H.P).
15. Performance security as mentioned in the document shall be obtained from the successful tenderer on the award of the contract in the shape of account payee demand draft or duly pledged fixed deposit receipts or bank guarantee from commercial bank will remain valid for one year from the date of completion of the contract if the supplier fails to repair the damaged/replaced within the one year guarantee/warrantee period, the performance security will be forfeited.
16. Refund of amount of security and earnest money deposited by the firm will be made after three months from the date of expiry of contract or after the date of satisfactory completion of the given supply orders. Keeping in view both these conditions while making applications for the refund of the amount of security. The concerned firm will have to furnish a certificate that the supply of all

the supply orders received by it has been made to the consignee according to the prescribed specifications to their entire satisfaction.

17. The earnest money of the bidders whose tender has not been approved/qualify would be return after proper scrutiny.
18. The tenders not confirming to the specifications/descriptions mentioned in bid schedule will be summarily rejected.
19. If the date of accepting/opening the tenders happens to be holiday, the tenders will be opened on the next day at the same time.
20. Telegraphic, Fax, Conditional and tenders without earnest money shall not be accepted.
21. Atal Bihari Vajpayee Govt. Institute of Engineering and Technology Pragatinagar Shimla, reserves the right to accept or reject any or all tenders without assigning any reason.
22. The bidders shall keep their offer open for 6 months from the date of opening the tenders. A bid valid for shorter period may be rejected by the purchaser as non-responsive.
23. The Director/Principal may appoint a "Negotiation Committee" if required.
24. The bidder will be responsible for damage or loss in transit and replace goods broken with in 10 days from the date of notice thereof.
25. GST rates shall be indicated separately otherwise rates will be deemed to be inclusive of such levies and taxes and no future increase in duty/taxes will be allowed.
26. Incase where it has not been specifically indicated in the schedule that only manufacturer can participate in tenders, in such cases the tenderers other than manufacturers should attach authorization letter from such principal manufacturer along with the tender.
27. The tender form along with the earnest money and forwarding letter on Letter Head Pad of the firm should be sent through Regd. Post/Speed Post/By Hand well in advance so as to reach to the office of Director-cum-Principal, ABVGIET Pragatinagar, Distt. Shimla-171202 (H.P) i.e **06/02/2026 up to 2:00PM**. The offer should be sent in a sealed envelope clearly indicating on the top the tender number, due date and the category of items.
28. All the bidders are required to sign each paper of the bid document along with the stamp of their respective organization. Without signature and stamp the tender submitted by them are likable for rejection.
29. The bidder has to attach the supporting documents/catalogues for all the items.
30. All the disputes shall be settled with in the jurisdiction of Sub-Judge Court, Shimla, Distt. Shimla H.P.
31. The earnest money deposited by the tenderers shall be forfeited in the following events:
 - i) A modification or withdrawal of tender after the deadline of submission of tenders and during the validity period.
 - ii) Refusal by the tenderer to accept an arithmetical error or otherwise appearing on the face of the tender.
 - iii) Failure on the part of the successful tenderer to provide performance security for the execution of the contract.
 - iv) Failure on the part of the successful tenderer to execute the contract as per terms and conditions of the tender.

32. The purchaser may, at its discretion, extend deadline for the submission of bids by amending the bidding documents, in which case all rights and obligations of the purchasers and bidders previously subject to the deadline will thereafter be subjected to the deadline as extended.
33. The purchaser has the right to ask the supplier(s) to produce the sample of the item(s) for the purpose of comparison of prices and to evaluate technically before placing supply order to the firm.
34. The purchaser or its representative shall have right to inspect or test the goods to confirm their conformity to the contract. The inspection and tests may be conducted on the premises of the supply at his costs.
35. The purchaser reserves the right to increase or decrease the quantity of goods to be procured.
36. An excused delay by the supplier in the performance of its delivery obligations shall render the supplier to any or all of the following:
 - i) Forfeiture of its bids security and its performance security.
 - ii) Imposition of liquidated damage.
 - iii) Termination of the contract for default.
37. If at any time during performance of the contract, supplier encounter conditions impending timely delivery of the goods and performance of the services, the supplier shall promptly notify the purchaser in writing of the facts of delay, its likely duration and its causes as soon as practicable after the receipt of the supplier notice. The purchaser shall evaluate the situation and may at its discretion extend the supplier time for the performance.

Special Conditions

1. All the items mentioned in the schedule shall be preferred from the single vendor.
2. The bidder must be a manufacturer or in supplying business as authorized distributor/stockiest for at least three years.
3. The bidder who is a potential authorized supplier/stockiest/dealer in the same business during last three years shall be preferred

I/We herewith enclose a sum of Rs. _____ as earnest money (EMD) in form of demand draft or equivalent and should I/We fail to execute an agreement embodying the above mentioned terms and conditions agree that the above sum as EMD shall be forfeited to the Director-cum-Principal, ABVGIET Pragatinagar, Distt. Shimla-171202 (H.P)

ALL THE ABOVE TERMS & CONDITIONS ARE ACCEPTABLE TO US

For (authorized Signatory)
Name of the Firm: _____
Complete Address: _____

Seal Stamp